

## The Inconsistency of Providing Ventilation with a T-piece Style Resuscitator: A Bench Model Study

Robin L. Angus BS, RRT-NPS  
*Lehigh Valley Health Network, Robin\_L.Angus@lvhn.org*

Kenneth Miller MEd, RRT-NPS  
*Lehigh Valley Health Network, Kenneth.Miller@lvhn.org*

Chris Fenstermaker MS, RRT-NPS  
*Lehigh Valley Health Network, Chris.Fenstermaker@lvhn.org*

Follow this and additional works at: <https://scholarlyworks.lvhn.org/patient-care-services-nursing>



Part of the [Equipment and Supplies Commons](#), [Medical Sciences Commons](#), and the [Nursing Commons](#)

## Let us know how access to this document benefits you

---

### Published In/Presented At

Angus, R., Miller, K., & Fenstermaker, C. (2013). *The inconsistency of providing ventilation with a T-piece style resuscitator: A bench model study*. Poster presentation.

This Poster is brought to you for free and open access by LVHN Scholarly Works. It has been accepted for inclusion in LVHN Scholarly Works by an authorized administrator. For more information, please contact [LibraryServices@lvhn.org](mailto:LibraryServices@lvhn.org).

# The Inconsistency of Providing Ventilation with a T-piece Style Resuscitator: A Bench Model Study

Robin Angus, BS, RRT-NPS; Kenneth Miller MEd, RRT-ACCs, RRT-NPS; Chris Fenstermaker, MS, RRT-NPS  
Lehigh Valley Health Network, Allentown, Pennsylvania

## Background:

- Providing consistent and safe ventilation is critical in optimizing gas exchange and guaranteeing patient safety
- Consistent and safe ventilation is most crucial in the neonatal patient population
- When a practitioner is unfamiliar with the neonatal patient population the above goals may become difficult to achieve

## T-Piece Resuscitator:

- Practitioner manually controls the delivered breath rate
- Practitioner has the ability to set a:
  - Peak Inspiratory Pressure (PIP)
  - Positive End Expiratory Pressure (PEEP)
  - FiO<sub>2</sub> when hooked up to a blender



## Study Design:

- To test the ability to provide a consistent and adequate ventilatory rate during a simulated neonatal resuscitation
- Goal:
  - Assess what percentage of practitioners would provide the recommended ventilatory rate of forty breaths per minute (two hundred breaths over 5 minutes) with a T-piece resuscitator
  - Evaluate the number of breaths delivered by each therapist over a five minute period
- Minutes 1, 2, & 5 were counted separately while minutes 3 & 4 were group together secondary to distractions to mimic chaos in a labor and delivery suite emergency.

## Methods:

- Seventy-one practitioners participated in the study
- The practitioners' years of experience and if they currently worked in an emergency department or labor and delivery patient population area was documented
- Practitioners received a simple review on how to properly utilize the T-Piece resuscitator that was provided
- A baby manikin head was used as the “patient” and was placed on a standard neonatal warmer
- Simulations were performed in two separate rooms with two separate counters who recorded the number of breaths delivered
- All set-ups were identical



Table 1. Title

	Labor & Delivery	Emergency Room	Not Specified
Participants	35	41	25
Percent Total	49.3%	57.7%	35.2%

Table 2. Title

Years of Experience	0-4	5-9	10-14	14+
Participants	4	8	9	50
Percent Total	5.6%	11.3%	12.7%	70.4%

## Results:

- Only 31 of the 71 (43.66%) practitioners met the desired breath rate of 40 breaths/minute (200 breaths over five the minute time frame)

Table 3. Title

	# Participants 38-42 BPM Range	% of Total Participants
1 Minute	10	14.0845
2 Minutes	14	19.7183
3 & 4 Minutes	18 (76-84 PPM range)	25.3521
5 Minutes	17	23.9437
	# Participants 180-220 BPM Range	
Total for 5 Minutes	31	43.6620

Table 4. Title

	100-149	150-199	200-249	250-299	300-349	350-399	400-449	450-499	500-549	550-600
Breaths	4	24	18	9	6	4	1	4	0	1

## Conclusion:

- Adequate ventilation is paramount in proving optimal gas exchange
- Inconsistent ventilation brings about the potential hazards of air trapping, hemodynamic insult, ventilation induced trauma, and inadequate gas exchange
- These adverse events may have long-term consequences in the neonatal patient population
- The study demonstrated that during a five minute simulated code practitioner delivered breaths tended to be inconsistent and vary from practitioner to practitioner
- Findings of this study warrants further research and exploration in the development of an automated rate delivery device

© 2013 Lehigh Valley Health Network