

Lehigh Valley Health Network
LVHN Scholarly Works

Patient Care Services / Nursing

Accuracy of Temporal Thermometer

Jennifer A. Hartman BSN, RN
Lehigh Valley Health Network, jennifer_a.hartman@lvhn.org

Sallie J. Rosenberger BSN, RN
Lehigh Valley Health Network, sallie_j.rosenberger@lvhn.org

Melissa L. Wessner BSN, RN
Lehigh Valley Health Network, melissa_l.wessner@lvhn.org

Follow this and additional works at: <https://scholarlyworks.lvhn.org/patient-care-services-nursing>



Part of the [Nursing Commons](#)

Let us know how access to this document benefits you

Published In/Presented At

Hartman, J., Rosenberger, S., & Wessner, M. (2016, August 19). *Accuracy of Temporal Thermometer*. Poster presented at LVHN Vizient/AACN Nurse Residency Program Graduation, Lehigh Valley Health Network, Allentown, PA.

This Poster is brought to you for free and open access by LVHN Scholarly Works. It has been accepted for inclusion in LVHN Scholarly Works by an authorized administrator. For more information, please contact LibraryServices@lvhn.org.

Accuracy of Temporal Thermometer

Jennifer Hartman, BSN, RN, Sallie Rosenberger, BSN, RN, Melissa Wessner, BSN, RN

Lehigh Valley Health Network, Allentown, Pennsylvania

BACKGROUND / INTRODUCTION

- Research on temperature readings shows core body temperatures, such as temporal artery (TA) thermometer readings, are more accurate than peripheral temperature readings such as those obtained from tympanic thermometers (Dunleavy, 2010).
- Purpose: to determine accuracy and consistency of body temperature readings with temporal artery thermometers in comparison to tympanic thermometers.

PICO QUESTION

- In the adult Medical-Surgical population, do temporal thermometers provide a more accurate and consistent measurement than tympanic thermometers?

EVIDENCE

- No discomfort for the patient (Allegaert, et. al, 2014).
- TA thermometers performed as better surrogate for brain temperature than tympanic thermometer (Bridges, et. al, 2009).
- Speed, easy to use, and lack of contact with body fluids (Dunleavy, 2010).
- Most reliable, less invasive way to measure temperature (Furlong, et. al, 2015).


METHODS

- Three experts (Nurse Residents) trained in use of the product.
- Obtained temporal and tympanic readings within 1 minute of each other over the course of 4 weeks.
- Temperature readings obtained and recorded every 4 hours.
- Tympanic readings converted to core body temperature by adding 0.8° F (Pompeii, 1999).
- Average deviations calculated from both tympanic and temporal artery thermometers.

OUTCOMES

- TA thermometer provides a more accurate and consistent measurement of body temperature than tympanic thermometers.
- TA thermometer are easy to clean and use.
 - Alcohol swab used to cleanse TA probe vs. single-use probe covers used for tympanic thermometers
- TA thermometers are more cost effective than tympanic thermometers.
 - Cost of each probe cover for tympanic thermometers: \$0.035
 - Cost of each alcohol swab to cleanse TA thermometers: \$0.01
 - Total cost of probes for FY '15: \$69,525.68
 - Total cost of alcohol swabs for FY '15: \$19,864.48

RESULTS



- Results from calculating deviations
 - Expert 1: tympanic 1.71 ° F – TA 1.13 ° F
 - Expert 2: tympanic 1.91 ° F – TA 1.03 ° F
 - Expert 3: tympanic 1.87 ° F – TA 1.44 ° F
 - Average Deviation of Thermometers:
 - Tympanic 1.83 ° F
 - Temporal 1.2 ° F

CONCLUSIONS

- There is less deviation of temperatures using the TA thermometer than the tympanic thermometer.
- The benefits of TA thermometers are:
 - Consistency in temperature readings
 - Cost effectiveness
 - Lifetime warranty



REFERENCES

- Allegaert, K., Casteels, K., Gorp, I., & Bogaert, G. (2014). Tympanic, infrared skin, and temporal artery scan thermometers compared with rectal measurement in children: A real life assessment. *Elsevier* 76. 34-38. doi: 10.1016/j.curtheres.2013.11.005
- Bridges, E., & Thomas, J. (2009). Noninvasive measurements of body temperature in critically ill patients. *Critical Care Nurse*, 29 (3), 94-97. doi:10.4037/ccn2009132
- Dunleavy, K. (2010). Which core body temperature measurement is most accurate? *Nursing*, 1. 18-19. doi: 10.1097/01.NURSE.0000390678.95642.7f
- Furlong, D., Carroll, D., Finn, C., Gay, D., Gryglik, C., & Donahue, V. (2015). Comparison of temporal to pulmonary artery temperature in febrile patients. *Dimension of Critical Care Nursing*, 34 (1). 47-52. doi: 10.1097/DCC.0000000000000090
- Pompeii, M. (1999). Temperature Assessment via the Temporal Artery: Validation of a New Method. Retrieved August 10, 2016 from <http://www.exergen.com/medical/PDFs/tempassess.pdf>.

© 2016 Lehigh Valley Health Network

A PASSION FOR BETTER MEDICINE.™

610-402-CARE LVHN.org